



Unique Products... Innovative Solutions!

CERTIFICATE OF CALIBRATION

This is to certify that all **TEST BLOCK KITS** are derived from an average of six readings taken with a Rex type D Durometer calibrated per ASTM D-2240. For periodic verification accuracy, take an average of three readings in different spots on each test block. The results should be approximate to those shown within ± 2 Duro points. Result comparisons may vary between durometers with non-similar foot configurations.

TEST BLOCK KIT S/N 083010MM6 **PE 141** was calibrated 08/30/10
using a Rex Digital Durometer DD-3-D S/N D-01047 (Calibrated 6/21/10 Due 6/21/11)
mounted in a Rex operating stand Model OS-1.
1 kg. Test results in test block kit. NIST# 822/266926-02

Work Performed By MICHELLE MANN

X 

Form No. 050206501P2

DUROMETER TEST BLOCK KIT

MADE IN USA

TYPE	BLUE 20	GRAY 40	BLACK 80
D	20.9	38.1	78.9

Result comparisons may vary between durometers with non-similar foot configurations.

SERIAL NO. 083010MM6

TEMPERATURE 70° HUMIDITY 45%

CALIBRATED

DATE DUE

The indicated readings are derived from an average of six reading taken with a Rex Durometer ($\frac{1}{2}$ " Diameter Foot) calibrated per ASTM D 2240.

For periodic verification of calibration accuracy, take an average of 3 readings in different spots on each test block. The results should be approximate to those shown above within ± 2 duro points (Type M ± 4 duro points).



Therma Products Innovative Solutions!