



Unique Products... Innovative Solutions!

## ***CERTIFICATE OF CALIBRATION***

This is to certify that all **TEST BLOCK KITS** are derived from an average of six readings taken with a Rex type A Durometer calibrated per ASTM D-2240. For periodic verification accuracy, take an average of three readings in different spots on each test block. The results should be approximate to those shown within  $\pm 2$  Duro points. Result comparisons may vary between durometers with non-similar foot configurations.

**TEST BLOCK KIT S/N** 083010MM **PE 142** was calibrated 08/30/10  
using a Rex Digital Durometer DD-3-A S/N A-0572 (Calibrated 6/21/10 Due 6/21/11 )  
mounted in a Rex operating stand Model OS-1.  
1 kg. Test results in test block kit. NIST# 822/266926-02

Work Performed By MICHELLE MANN

X A handwritten signature in blue ink, appearing to read "Michelle Mann", written over a horizontal line.

Form No. 050206501P2

# DUROMETER TEST BLOCK KIT

MADE IN USA

TYPE	BLUE 30	RED 40	ORANGE 50	YELLOW 60	BROWN 70	BLACK 80	GRAY 90
A	30.5	43.0	52.4	63.2	73.0	79.0	92.4

*Result comparisons may vary between durometers with non-similar foot configurations.*

SERIAL NO. 083010MM

TEMPERATURE 70° HUMIDITY 45%

CALIBRATED

DATE DUE

The indicated readings are derived from an average of six readings taken with a Rex Durometer ( $\frac{1}{2}$ " Diameter Foot) calibrated per **ASTM D 2240**.

For periodic verification of calibration accuracy, take an average of 3 readings in different spots on each test block. The results should be approximate to those shown above within  $\pm 2$  duro points (Type M  $\pm 4$  duro points).

